

FAST Sphyg

Aneroid Sphygmomanometer

The FAST Sphyg is an aneroid sphygmomanometer intended to be used with a stethoscope or Doppler for indirect measurement of arterial blood pressure.

- **Push button inflation**
Reduces hand and wrist strain!
- **User controlled deflation**
For accurate readings!
- **Ergonomic design**
For user comfort!
- **Rechargeable battery**
- **For use with a wide variety of cuff sizes***

PRODUCT SPECIFICATIONS:

Pressure Range:	0 - 300 mmHg
Pressure Accuracy:	$\leq \pm 3$ mmHg
Max. Air Leakage Rate:	≤ 4 mmHg / minute
Power:	Rechargeable Li-ion Battery
Voltage:	3.7 Volts
Battery Life (before recharge):	Approx. 250 readings
Inflation Speed:	Variable
AC Adaptor:	
Input Voltage:	100 to 240 Vac.
Output Voltage:	DC5V, 1.2A
Environmental Temperature Range:	
Operation:	10° to 40° C
Storage/Transport:	-20° to 55° C
Relative Humidity Range:	
Operation:	15 to 85% (non-condensing)
Storage/Transport:	90%
Dimensions:	216 (L) x 76 (W) x 89 (H) mm
Dial Face:	49 mm (Diam.)
Weight:	310 grams
Electrical safety:	Conforms to IEC60601-1 Class IIa equipment Internally powered equipment Type B applied part Conforms to Directive 2011/65/EU on the restriction of the use of certain hazardous substances in electrical and electronic equipment Retains calibration for drops under 25 cm
Shock proof:	
Specifications are subject to change	

INCLUDES: FAST Sphyg, (1) AC charger, (1) wrist strap, (1) guard

All cuffs sold separately

VASCULAR INSTRUMENTS
KOVEN
TECHNOLOGY INCORPORATED

Koven Technology, Inc.
Your Vascular Healthcare Partner

Koven Technology, Inc.
477 N. Lindbergh Blvd., Suite 220, St. Louis, MO 63141
Tel: 800-521-8342 / 314-542-2101 • Fax: 314-542-6020
info@koven.com • www.koven.com



 **Patent Pending**

GTIN: 00894912002920

** Contraindication: Safety and effectiveness with neonate cuff sizes 1 through 5 is not established.*

Caution: Federal (USA) law restricts this device to sale by or on the order of a physician. Rx only.

Caution: The FAST Sphyg has not been evaluated for safety and compatibility in the MR environment. The FAST Sphyg has not been tested for heating or migration in the MR environment.

For a complete listing of indications, contraindications, cautions and warnings, please refer to the operation manual provided with this product.

GLN: 0894912002005

© Koven Technology, Inc. DS-1604 rev. 3